

CERTIFICATE OF CONFORMITY

Order Code: **232-041**

Therma 22 Plus Thermometer

Type 'T' Thermocouple or Thermistor Probes

Range: 'T' T/C: -199.9 to 400°C, Therm: -39.9 to 149.9°C

Resolution: 0.1°C to 300°C thereafter 1°C

Accuracy: 'T' T/C: $\pm 0.5^\circ\text{C}$, -50 to 150°C

Thermistor: $\pm 0.4^\circ\text{C}$, -25 to 110°C

Instrument only accuracy $\pm 0.2^\circ\text{C}$

Made in Britain

Conforms to RoHS & REACH Legislation

Conforms to EC Reg 1935/2004 & EU Reg 10/2011

DECLARATION OF CONFORMITY

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

We also confirm that when used in an appropriate manner, the instruments are suitable for use within the Food Industry, and also, where appropriate and required, comply with EEC Directives with regard to European Legislation and CE Marking requirements.

Including EC Regulation No 1935/2004 on Materials and articles intended to come into contact with food; & Commission Regulation (EU) No 10/2011 on plastic materials and articles intended to come into contact with food.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd



Andy Reid

Quality Manager

Valid as of date of purchase.

Easting Close
Worthing
West Sussex
BN14 8HQ
T +44 (0)1903 202151
E sales@etiltd.com
W etiltd.com