

InstaTest[®] ANALYTIC

DRINKING WATER 6

Nitrate, Nitrite, pH, Free Chlorine, Total Chlorine & Total Hardness



25 Test Strips

Code 2933-G


YalITech[®]
Yalitech Instruments
San Diego de Chile, Rio Rodrigo 564, Parque de Negocios ERGA, Providencia
www.yalitech.cl | ventas@yalitech.cl | +56 21 2598221

Immerse strip for 2 seconds.

Remove with pads face up.

SHAKE ONCE TO REMOVE

EXCESS WATER. Immediately

read **Free Chlorine, Total**

Chlorine, Total Hardness and

pH. Wait 30 seconds, then

read: **Nitrite > Nitrate.**

- Keep wet fingers out of the vial
- Close vial tightly after removing strip
- Store in a cool, dry place

LaMotte Company
Chestertown, Maryland 21620 USA
www.lamotte.com



Free Chlorine ppm	0	0.5	1	3	5	10
Total Chlorine ppm	0	0.5	1	3	5	10
Total Hardness ppm (ppg)	0 (0)	50 (3)	100 (5.8)	200 (11.7)	400 (23)	
pH	4	5	6	7	8	10
Nitrite as NO ₂ -N ppm	0	0.5	1	5	10	
Nitrate as NO ₃ -N ppm	0	5	10	25	50	