



25 Pool & Spa
Test Strips

25 Tiras de análisis
para piscinas y spas

25 Languettes d'analyse
pour piscines et spas

NITRATE
NITRATE
NITRATO



Yalitech Instruments

Santiago de Chile, 810 Refugio 9648, Parque de Negocios ENEA, Pudahuel.
www.yalitech.cl - ventas@yalitech.cl - (+56 2) 28988221

Code 3012-G
Made in USA

Nitrate, Nitrite, Nitrito ppm	0	10	30	60	120	200

Sumerja la tira en la muestra durante 2 segundos. Retírela con el lado coloreado hacia arriba. NO SACUDA EL EXCESO DE AGUA.. Mantenga la tira horizontal a la sombra por **60 segundos** y haga la lectura.

No toque el envase con los dedos húmedos.

Cierre bien el envase luego de sacar la tira.

Guarde el envase en un lugar frío, seco y oscuro.

Immerger pour 2 secondes dans l'échantillon d'eau. Retirer la bande face vers le haut. NE PAS SECOUER. À l'ombre, garder le niveau de la bande pour **60 secondes** et lire.

Ne pas introduire les doigts trempés dans la bouteille.

Refermer le capuchon.

Entreposier dans un endroit frais, sec et ombragé.

Dip pad into sample and immerse for 2 seconds. Remove strip with pad face up. DO NOT SHAKE OFF EXCESS WATER. In shade, hold pad level for **60 seconds** and read the strip.

Keep wet fingers out of the bottle.

Close vial tightly after removing strip.

Store indoors in a cool, dry, dark place.

LaMotte Co., Chestertown, MD 21620 USA • www.lamotte.com/insta



6 37395 45337 3